

Standard Operating Procedure
(Laboratory Practices)

Infectious Agent and Strain: Unfixed Human Primary Cells

Lab/Location(s): Mayer 307

Principal Investigator: Suzan Lazo

Protocol Name & Number: Sorting of Unfixed Human Primary Cells

Current Revision Date: 04/01/2025

Description of Possible Risks:

The Facility will comply with BSL2 Enhanced Safety practices.

Unfixed Human Primary Cells (HPC) Sorting Procedure:

HPC will be sorted on BD FACSAria III SORP, BD Symphony S6 or Sony MA900 cell sorters housed within the Biocontainment hoods. This SOP clearly addresses the procedures for evacuating the sort chamber of aerosols prior to opening the sort chamber, including a stated time to wait after a clog induced stream deviation. This is necessary as most cell sorters do not possess an interlock designed to prevent the operator from opening the sort chamber after a nozzle obstruction with subsequent stream deviation.

Prior to Sorting:

1. Fill Sheath tank
2. Add Bleach to the waste tank to ensure a final 10% solution post sort waste
3. Fill Spray bottle with fresh 10% Bleach solution
4. Inspect sort chamber, nozzle, o-ring, and deflection places for saline build up and clean if necessary.
5. Researchers must use universal safety precautions, wear lab coats and gloves

Procedures during Sorting:

1. Researchers must use universal safety precautions, wear lab coats and gloves.
2. Biocontainment hood must have the glass doors closed. Once sorting has started access is permitted through the sash opening.
3. Biocontainment hood blower switch must be on. This will engage the Air Pressure Alarm used to indicate if there is a high or low air flow condition present for more than 3 seconds.
4. Samples must be filtered by researcher and solid caps placed on tubes immediately prior to sorting to minimize clogs and ice buckets must be placed on the table inside of biocontainment hood.

5. Sample tubes should be filled with maximum amount of sample to minimize loading and unloading samples.
6. Enable "sweet spot" to ensure stream stability.
7. Install correct collection receptacle, open sort collection chamber and align sort streams
8. Close sort collection chamber door prior to starting sort
9. Solid (not blue filter cap) test tube caps must remain on the samples.
10. Samples may only be uncapped immediately prior to placement on the sample tube holder
11. Collection tubes must be capped prior to removal from the sort collection chamber
12. Collections tubes must be wiped down with 10% bleach prior to removal from the sort collection chamber
13. Following the sort, the operator must disinfect the sample tubing.
 - a. Run 5 minutes of 10% Bleach
 - b. Run 3 minutes of ddH₂O
14. Disinfecting tubes must be capped and disposed of in the biohazard waste bag

Nozzle Obstruction cleaning procedure:

In the event the HPC cell population clogs the sorter (tubing or nozzle):

1. Operator will stop the sort stream: turn off the stream using the button labeled with a '✓' on the Breakoff window. This will shut off the stream, unload the sample and close the aspirator door. Alternatively, pressing the Large Red Emergency Stop Button to the right of the sample stage will also stop the sheath and sample stream.
2. Re-cap the sample tube and wipe the exterior of the tube with 10% bleach and place in ice bucket
3. Open aspirator drawer using software controls
4. Evacuate the collection chamber for 60 seconds; increase Aerosol Management System (AMS) evacuation rate to 100% vacuum. MUST WAIT 60 SECONDS
5. Remove and re-cap collection tubes and wipe the exterior of tube with 10% bleach and place in ice bucket
6. Close aspirator drawer using software controls.
7. Operator will then perform the BD Clean Flow Cell Procedure with ddH₂O 3x
 - a. Open aspirator drawer using software controls
 - b. Evacuate collection chamber for 60 seconds
 - c. Place ddH₂O on sample station
 - d. Diva Software: Select from the menu - Instrument > Cleaning Modes > Clean Flow Cell.
 - e. Repeat twice
8. Restart the sort stream

9. In the event the nozzle obstruction remains as evidenced by an incorrect droplet image formation, the Operator must open aspirator drawer wait 3 minutes to allow aerosols in the sort chamber to be evacuated.
10. Operator will decontaminate the sort chamber with 10% bleach spray and Super Sani Cloths. With stream turned off, open the sort block chamber door and dry plates and surfaces as needed.
11. Remove the nozzle; wipe nozzle with Super Sani-cloth wipes , place the nozzle in ddH₂O for in a 15cc conical tube with 1ml ddH₂O, cap and ultrasonicate for 60 seconds.
12. Set AMS unit to 20% vacuum, replace nozzle, and ensure all chamber doors are closed then restart sort stream, verify that correct droplet stream image is present.
13. All tubes and wipes must be disposed of in biohazard bag.

Post-Sorting procedure: Decontamination Procedure:

After cell sorting is complete the operator is responsible for disinfecting the areas used for HPC cell sorting.

Disinfect sample lines using a freshly made 10% bleach solution as follows:

- a. Fill a tube with a volume of 10% bleach equal to or greater than the volume of sample that was sorted and place on the sample stage.
- c. Run 10% Bleach for 5 minutes at maximum flow rate
- d. Run DI water for 3 minutes at maximum flow rate

The following areas must be disinfected with 10% Bleach and the Super Sani-cloth wipes to clean accordingly:

- Sort chamber
- Sample holder
- Immediate surrounding surfaces
- Sort Chamber
- Sample Station doors
- Computer workstation area and keyboard

Contact the Flow Cytometry Facility (617-632-3179)
Suzan Lazo (617-632- 2049) for more information.